

## **TMDE Batteries**

The MMOD-CA requests that you remove the 9-volt battery(s) in test, measurement and diagnostic equipment (TMDE) when it is not in use. They are receiving much of the TMDE with corrosion damage because old batteries were left inside.

POC: MMOD-Tracy, CA phone: 209-839-4556 or DSN: 462-4556