Repairer Preventive Maintenance Checks and Services

4110-01-159-6922
Refrigerator, Mechanical, Blood Bank, Model 139875

[M-Monthly, Q-Quarterly, S-Semiannually, and A-Annually]

ITEM NO	INTERVAL	ITEM TO BE INSPECTED AND PROCEDURE	IS NOT MISSION CAPABLE IF:
1	S	**Refrigerator** a. Ensure that a copy of the manufacturer's manual is on hand.	
		b. Inspect the refrigerator for obvious signs of damage such as cracks, dents, leaks or broken components.	The power cord is cracked or frayed, wires are not covered by the cord insulation, or the damage prevents the refrigerator from operating.
2	A	**Installation of the Refrigerator** Verify that the refrigerator has been installed according the Operator Preventive Maintenance Checks and Services.	
3	A	**Maintaining Refrigerator** a. Verify that the refrigerator has been maintained according to the Operator Preventive Maintenance Checks and Services.	
		b. Verify electrical safety.	The refrigerator fails any of the electrical safety tests.