LARYNGOSCOPE SET

MANUFACTURER: WELCH ALLYN
Manufacturer Model# MIL-5062
Manufacturer's PHONE: (315) 685-4560 FAX: (315) 683-3361
Manufacturer's Website: <http://www.welchallyn.com>

AAC 'J' ITEM NSN 6515-01-526-4240

UAs: 267B , 268B , 269B , 623A , P301 , P307 , P308

USAMMA POC: Customer Relationship Management 301-619-1288/4301 - DSN 343-1288/4301
UNIT PRICE: $1,926.06

System Description: Laryngoscope Set Fiber Optic Battery Operated with Case.

Features:

- Reduced handle and blade weight for improved balance and maneuverability facilitates even the most difficult intubations.
- Repairable, allowing replacement of deteriorating light pathways for a fraction of the cost of a replacement blade.
- Improved design with reduced proximal blade height assists in intubating patients with limited oral access.
- Single-piece stainless steel blades ensure blade integrity (no soldered joints).

LARYNGOSCOPE SET COMPREHENSIVE
PART # MIL-5062
NSN 6515015264240



WelchAllyn®

Part No. 690048-1 Rev D

4341 State Street Road
PO Box 220
Skaneateles Falls, NY 13153-0220 USA
Tel: +1 315 685 4560
Fax: +1 315 685 3361
www.welchallyn.com

1

Fiber Optic Laryngoscope Blades and Handles

Fiber Optic Laryngoscope Blades

Care and Maintenance

Welch Allyn laryngoscope blades are crafted from enduring stainless steel. To insure maximum life and performance the following instructions should be strictly adhered to.

Miller, MacIntosh, and E-MacIntosh styles of fiber optic blades all have removable light pipes. This unique feature eliminates the expense of replacing the entire blade due to the fiber bundle deterioration that results from repeated exposure to disinfection and sterilization agents. Now the bundle can simply be replaced at a fraction of the cost of a new blade.

Cleaning Procedure

Immediately after use, blades should be rinsed in clean tap water to remove any residue. The fiber optic light pipe should be removed from the blade following the procedure outlined in the Fiber Optic Light Pipe Replacement section. Both the light pipe and blade should then be gently scrubbed in soapy water, with a soft brush, to provide a thorough physical cleaning.

All Welch Allyn blades are compatible with enzymatic cleaners. Refer to manufacturer's instructions for recommended exposure times and solution strengths.

After cleaning, thoroughly rinse the light pipe and blade, dry, and reassemble (see Fiber Optic Light Pipe Replacement section) prior to disinfection or sterilization.

Note If germicidal solutions are required for cleaning, please contact Welch Allyn for compatibility.

WARNING Ultrasonic cleaning is not recommended.

LITERATURE OPERATION
PART # 690048-1
NSN 7610015437003

PELICAN CASE, CASE IS ONLY AVAILABLE FOR INITIAL PURCHASE

SPARE LAMPS # 06000-U NSN 6515-01-519-4257
LAMP # 608125-501 NSN 6240-01-526-9631
MILLER SIZE 0 # 68060 NSN 6515-01-527-2281
MILLER SIZE 1 # 68061 NSN 6515-01-442-3002

STUBBY HANDLE # 60815 NSN 6515-01-527-2289
“AA” HANDLE # 60814 NSN 6515-01-445-1081
“C” SIZE HANDLE # 60813 NSN 6515-01-445-1082

ENGLISH MACINTOSH SIZE 1 # 69211 NSN 6515-01-467-3729
ENGLISH MACINTOSH SIZE 2 # 69212 NSN 6515-01-467-4411
ENGLISH MACINTOSH SIZE 3 # 69213 NSN 6515-01-467-3430
ENGLISH MACINTOSH SIZE 4 # 69214 NSN 6515-01-467-3431

MILLER BLADE SIZE 2 # 68062 NSN 6515-01-527-2283
MILLER BLADE SIZE 3 # 68063 NSN 6515-01-519-4239
MILLER BLADE SIZE 4 # 68062 NSN 6515-01-527-2286

LARYNGOSCOPE SET
PHYSICAL & ENVIRONMENTAL SPECIFICATIONS

Item Dimensions	**Height (Inches)**	**19.5**
	Width (Inches)	**14**
	Length (Inches)	**4.75**
	Cube (Cubic Feet)	**0.7504**
	Weight	**9.545 lbs**
Ship Dimensions	**Height (Inches)**	
	Width (Inches)	
	Length (Inches)	
	Cube (Cubic Feet)	
	Weight	
Storage Temp		
Operating Temp		
Storage Humidity		
Operating Humidity		
Amps		
Watts		
Power Requirements	**60813 requires 2 standard C size batteries; 60814 and 60815 require 2 standard AA size batteries**	
Notes		

LARYNGOSCOPE SET
ACCESSORIES & CONSUMABLES - STARTUP

NSN	Notes	Nomenclature	MFR PN Welch Allyn	DIST PN	QTY	UI	Unit Price	Total Price	Shelf Life Mon.
6515015264240		Laryngoscope Set	MIL-5062		1	EA	$1,926.06	$1,926.06	N/A
6240015269631		Cartridges, Lamp	608125-501		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6240015269660		Pen Lite	76600		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515014423002		Miller Blade Size 1	68061		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515014451082		Handle Laryngoscope Size C Fiber Optic	60813		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515014673429		English Macintosh Size 1	69211		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515014673430		English Macintosh Size 3	69213		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515014673431		English Macintosh Size 4	69214		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515014674411		English Macintosh Size 2	69212		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015194239		Miller Blade Size 3	68063		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015194257		Light Bulb Laryngoscope	06000-U		4	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015194259		Handle Laryngoscope Size AA Fiber Optic	60814		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015272281		Miller Blade Size 0	68060		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015272283		Miller Blade Size 2	68062		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015272286		Miller Blade Size 4	68064		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
6515015272289		Stubby Handle	60815		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
7610015437003		Literature Operator	690048-1		1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A
Not Required	Not Available After Initial Purchase	Pelican Case			1	EA	Incl w/NSN 4240	Incl w/NSN 4240	N/A